

BASIC LS-APGD SOURCE OPERATION




Fig. 1b

PROPOSED IMPLEMENTATION OF LS-APGD WITH MICROFLUIDIC DEVICES




Fig. 1c




Fig. 2a






Fig. 2b









Fig. 4a







Fig. 5






Fig. 7





Fig. 8





REPLACEMENT SHEET

ANALYTICAL RESPONSE FUNCTIONS AND LIMITS OF DETECTION FOR THE LS-APGD DEVICE. SOLUTION FLOW RATE = 1 mL/MIN., ELECTROLYTE pH = 1, INTER ELECTRODE GAP = 1 mm, INJECTION VOLUME = 5μ L.

ELEMENT	ELEMENT WAVELENGTH (nm)	PEAK HEIGHT EQN. R ²	PEAK AREA EQN.	LOD LOD
Na	589.0	Y=0.421x + 42.8 0.9859	Y=0.421x + 42.8 0.9859 Y=15.81x + 978.6 0.9784	12 (60)
Fe	248.3	Y=1.06x - 102.1 0.9365 Y=45.80x - 6649 0.909	Y=45.80x - 6649 0.909	12 (60)
Pb	405.8	Y=1.18x - 10.45 0.977	Y=1.18x - 10.45 0.977 Y=16.16x - 419.7 0.9298	14 (70)

Fig. 10








Fig. 13




Fig. 14

Selenoamino Acid Separation




Fig. 15